



RapidChek® *Campylobacter*

RapidChek® *Campylobacter* Test Kit, 10005433

Test Kit Includes:

- 50 test strips in a desiccated canister
- 50 transfer pipettes
- 50 sample tubes
- Package Insert

Does not include media.

RapidChek® *Campylobacter* Media

- 500 g – 10006043
- 5 kg – 10006042

RapidChek® *Campylobacter* Strips, 10006045

Test Kit Includes:

- 10 x 50 test strips in desiccated canisters
- Does not include tubes, pipettes, or media.

Also Available:

- RapidChek® SELECT™ *Salmonella*
- RapidChek® SELECT™ *Salmonella* Enteritidis
- RapidChek® CONFIRM™ *Salmonella* Enteritidis IMS
- RapidChek® *Listeria*
- RapidChek® *Listeria* NextDay™
- RapidChek® *E. coli* O157 (including H7)
- RapidChek® CONFIRM™ non-O157 STEC IMS





1 Media Preparation

Add media to sterile water.
DO NOT AUTO-CLAVE.



Sterile Water (Liters)	RapidChek® <i>Campylobacter</i> Media (Grams)
1	55.2

Rehydrated Media Shelf-Life
6 hours at room temperature.

2 Sample Enrichment

Add prepared RapidChek® *Campylobacter* Media to the sample. Incubate at 42 °C in a vented culture flask.



Sample Type	Volume of Media to add (mL)	Incubation Time (Hours)
30 mL carcass rinse	30 mL	47 – 49
30 mL ground chicken slurry	30 mL	47 – 49
25 mL turkey carcass rinse	25 mL	47 – 49

3 Transfer

Transfer 0.4 mL of the enriched sample into a supplied tube.



4 Test

Place strips into tubes with arrows facing down.



5 Read Result

After 20 minutes, read result.
1 line = Negative
2 lines = Positive

